



TENSIOMETRE AUTOMATIQUE POIGNET

MESURE PRESSION ARTERIELLE DIASTOLIQUE ET SYSTOLIQUE, ET MESURE DU POULS AU POIGNET

- + Garantie 2 ans
- + Mesure facile (1 seul bouton)
- + Mesure par la méthode oscillométrique
- + Détection des arythmies cardiaques
- + Validé cliniquement
- + Écran large
- + Manchon taille large (13,5 à 21,5 cm)
- + 2 utilisateurs x 199 Mémoires de mesure par utilisateur
- + Fonction date et heure incluse pour les mémoires
- + Calcul de moyenne de 3 mesures
- + Arrêt automatique et économie d'énergie
- + Plages de mesures
 - Dia. / Sys. : 0 - 299 mmHg
 - Puls. : 40 - 199 puls./min
- + Précision de mesures
 - Dia. / Sys. : ± 3 mmHg
 - Puls. : $\pm 5\%$
- + Fourniture de 2 piles AAA incluses
- + CE 0123

 Détection arythmie cardiaque Detectie van hartritmestoornissen Erkennung von Herzrhythmusstörungen Cardiac arrhythmia detection	 2x199 mémoires 2x199 sets herinneringen 2x199 Sätze von Erinnerungen 2x199 sets memory	 Écran large Breedbeeld Breitbild Wide screen	 Taille du manchon Manchet maat Manschettengröße Cuff size (13,5 – 21,5 cm)
 Classification OMS WHO-classificatie WHO-Klassifizierung WHO classification	 3 Mesures moyennes 3 Metingen gemiddeld 3 Messungen Durchschnitt 3 Times average	 Fonction date / heure Datum / tijd functie Datum / Zeitfunktion Date / time function	

Désignation	EAN13	CNK	Colisage préconisé
OFFISOINS Tensiomètre automatique poignet	7863124802331	4613519	1



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)

(Devices in Class IIa, IIb or III)

No. G1 078672 0014 Rev. 01

Manufacturer:

Shenzhen Urion Technology Co., Ltd.

Floor 4-6th of Building D

Jiale Science & Technology Industrial Zone

No.3, ChuangWei Road

Heshuikou Community, MaTian Street

GuangMing New District

518106 Shenzhen

PEOPLE'S REPUBLIC OF CHINA

Product Category(ies): Digital Blood Pressure Monitors, Infrared Thermometer, Compressor Nebulizers

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G10786720014Rev.01

Report No.:

GZ2117902

Valid from:

2021-04-12

Valid until:

2024-05-26

Date,

2021-04-12

Christoph Dicks

Head of Certification/Notified Body